



EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 102326 0002 Rev. 01

Manufacturer:

Thora Tech GmbH

Gutfleischstr. 3-5
35390 Gießen
GERMANY

SRN Manufacturer - DE-MF-000005161

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s). The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result.

The quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis.

The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. The surveillance assessment shall also include an assessment of the technical documentation for the device or devices concerned on the basis of further representative samples. All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with.

For details and certificate validity see: [www.tuvsud.com/ps-cert?q=cert:G10 102326 0002 Rev. 01](http://www.tuvsud.com/ps-cert?q=cert:G10_102326_0002_Rev.01)

Report No.:

713369417

Preceding Certificate No.:

G10 102326 0002 Rev. 00

Valid from:

2025-11-21

Valid until:

2029-11-27

Date of Initial Issuance:

2024-11-28

Christoph Dicks

Head of Certification/Notified Body

Issue date: 2025-11-21



EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 102326 0002 Rev. 01

Classification: Class IIb
Device Group: Z120301 - ANAESTHESIA AND PULMONARY VENTILATION SUPPORT INSTRUMENTS
Intended Purpose: The universal jet ventilator monsoon 4 is designed for jet ventilation of patients in the operating theatre during laryngoscopy, rigid bronchoscopy or microsurgery with or without a laser. The aim of jet ventilation is to secure patients oxygenation using compressed air, medical oxygen or a combination of both. The minimum body weight for using the monsoon 4 is 3 kg

Classification: Class IIa
Device Group: R010199 - NASO-OROPHARYNGEAL TUBES - OTHER
Intended Purpose: -

The validity of this certificate depends on conditions and/or is limited to the following: ./.

Revision History:

Rev.	Dated	Report	Description
00	2024-11-28	713227351	Initial issuance
01	2025-11-21	713369417	Supplemented: Device(s)/group of device(s) added